


National Institute
of Standards and Technology



National Voluntary
Laboratory Accreditation Program

ISO/IEC 17025:1999
ISO 9002:1994

Scope of Accreditation



Page: 1 of 1

Revised Scope 08/14/2003

IONIZING RADIATION DOSIMETRY

NVLAP LAB CODE 100539-0

U.S. ARMY RADIATION STANDARDS & DOSIMETRY LABORATORY

Attn: AMSAM-TMD-SR-D, Bldg. 5417

Redstone Arsenal, AL 35898-5000

Mr. Patrick Kuykendall

Phone: 256-876-3340 Fax: 256-955-6413

E-Mail: patrick.kuykendall@redstone.army.mil

Scope of Accreditation:

This facility has been evaluated and deemed competent to process the radiation dosimeter listed below through employing a Panasonic Model 710 reader.

This facility is accredited to process the following dosimeter by virtue of actual demonstration of compliance with ANSI HPS N13.11-2001 through testing.

Panasonic Model UD802AS in a Panasonic UD-874AT holder for ANSI-N13.11-2001 categories IA, IB, IC, IIA, IIB, IIC, IID, IIIA, IIIB, IIIC, IVA, IVB, IVC, VAB, VICB.

December 31, 2003

Effective through

A handwritten signature in black ink, appearing to read "John P. Waples".

For the National Institute of Standards and Technology